

Eljintek, Inc.
Ms. Lydia Zhang
-
6F-1, No.286-3, Shin Ya Road
Chien Chen Dist.
Kaohsiung 806
Taiwan

Date : 23.03.2018
Our ref. : ZTW1
Your ref.: GW-26012018-MW/Co-li

Ref : TA License

Type of Equipment : AC/DC MEDICAL ADAPTOR
Model Designation : See Certificate
Certificate No. : TA 50402959 0001
Report No. : 10042232 005

Dear Ms. Lydia Zhang,

Enclosed please find above certification documents.
Please forward these originals to the certificate holder.

If you contact our office, please quote our reference above.

We thank you for your cooperation.

the certificate holder is: MEAN WELL Enterprises Co., Ltd.

With kind regards,

Certification Body



Dipl.-Ing. (FH) A. Klinker

Enclosure

TÜV RHEINLAND TAIWAN LTD.

11F., No. 758, Sec. 4, Bade Rd.,
Songshan Dist., Taipei City 105,
Taiwan R. O. C.
Tel. (02) 2172-7000
Fax (02) 2528-0018
<http://www.tuv.com>

TAICHUNG BRANCH:
No. 9, Lane 36, Sec. 3, Minsheng Rd.,
Daya Dist., Taichung City 428
Taiwan, R. O. C.
Tel. (04) 2560-2998
Fax (04) 2566-3598

MEAN WELL Enterprises Co., Ltd.
Mr. Frank Lin Safety Engineer
-
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Date : 23.03.2018
Our ref. : ZTW1
Your ref.: GW-26012018-MW/Co-li

Ref : TA License

Type of Equipment : AC/DC MEDICAL ADAPTOR
Model Designation : See Certificate
Certificate No. : TA 50402959 0001
Report No. : 10042232 005

Dear Mr. Frank Lin,

The above specified technical equipment has been tested and found to be in accordance with the relevant requirements.

Enclosed you will find the certificate of approval.

If cancellation of the certificate is submitted by November 15 in a given year, no fee will be charged for the following year.

We appreciate your kind support and would like to offer our assistance and continuous services in the future.

Please contact our office for approval of your new products.

With kind regards,

Certification Body



Dipl.-Ing. (FH) A. Klinker

cc: Eljintek, Inc.

Enclosure

TÜV RHEINLAND TAIWAN LTD.

11F., No. 758, Sec. 4, Bade Rd.,
Songshan Dist., Taipei City 105,
Taiwan R. O. C.
Tel. (02) 2172-7000
Fax (02) 2528-0018
<http://www.tuv.com>

TAICHUNG BRANCH:
No. 9, Lane 36, Sec. 3, Minsheng Rd.,
Daya Dist., Taichung City 428
Taiwan, R. O. C.
Tel. (04) 2560-2998
Fax (04) 2566-3598

Certificate



Certificate no.

TA 50402959 01

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuguan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

030-0000334704

Test report no.:

ZTW1- 10042232 005

Client Reference: GW-26012018-MW/Co-li

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-6:2010+A1
EN 60601-1-6:2010+A1
IEC 62366:2007+A1
EN 62366:2008+A1

Certified Product:

(AC/DC MEDICAL ADAPTOR)

License Fee - Units

Type Designation	: GEM18xyzzzzz (MW)	5
x stands for	: B, E, I	1
y stands for	: 05, 07, 09, 12, 15, 18, 24, 28, 48	1
z stands for	: 0-9, A-Z, "-" or blank	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 0.45-0.2A	
Output Voltage	: see Appendix	
Output Current	: see Appendix	
max. Power Output	: see Appendix	
max. Ambient Temperature:	40°C	
Protection Class	: II	

This certificate is valid for maximum 5 years.

Appendix: 1

8

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Date of Issue

(day/mo/yr)

23/03/2018



Dipl.-Ing. (FH) A. Klinker

Appendix to TÜV approved Certificate No.: TA 50402959

Certified Product : AC/DC MEDICAL ADAPTOR
Report Number : 10042232 005
Type Designation : GEM18xyzzzzz
 (x = B, E, I; y = 05, 07, 09, 12, 15, 18, 24, 28, 48;
 zzz = 0-9, A-Z, "-" or blank)

Co-Model	Output Voltage (Vdc)	Output Current (A)	Max Power
GEM18x05zzzzz	5 - 6	3.00 - 2.50	15
GEM18x07zzzzz	6 - 8	2.50 - 1.87	15
GEM18x09zzzzz	8 - 11	2.25 - 1.64	18.04
GEM18x12zzzzz	11 - 13	1.64 - 1.38	18.04
GEM18x15zzzzz	13 - 16	1.38 - 1.13	18.08
GEM18x18zzzzz	16 - 21	1130 - 860mA	18.08
GEM18x24zzzzz	21 - 27	860 - 670mA	18.09
GEM18x28zzzzz	27 - 33	670 - 550mA	18.15
GEM18x48zzzzz	33 - 58	550 - 320mA	18.56

Date: 23rd March, 2018



Certification Body

Dipl.-Ing. (FH) A. Klinker