



## Declaration of Conformity

For the following equipment :

Product Name: Medical Type Switching Power Supply

Model Designation: NMP1K2-aaaaaa-xx(a=C , E , H , K or # : x=0~9) C=NMS-240-05 , E=NMS-240-12 , H=NMS-240-24 , K=NMS-240-48,#=Blank

is herewith confirmed to comply with the requirements set out in the Council Directive, the following standards were applied :

**RoHS Directive (2011/65/EU), (EU)2015/863**

**MDR Directive (EU) 2017/745**

EN 60601-1:2006+A11+A1+A12

TUV certificate No : TA50393544 (for NMP1K2)

TUV certificate No : TA50393494 (for NMS-240)

### EMI (Electro-Magnetic Interference)

Conducted emission / Radiated emission

EN 55011:2016+A1:2017  
EN 55032:2015

Class B

Harmonic current EN 61000-3-2:2014

Voltage flicker EN 61000-3-3:2013

### EMS (Electro-Magnetic Susceptibility)

EN 60601-1-2:2015 EN 55024:2010+A1:2015

ESD air EN 61000-4-2:2009 Level 4 15KV

ESD contact EN 61000-4-2:2009 Level 4 8KV

RF field susceptibility EN 61000-4-3:2006+A1:2008+A2:2010 Level 3 10V/m(80MHz-2.7GHz)

RF field susceptibility EN 61000-4-3:2006+A1:2008+A2:2010 Table 9 9~28V/m (385MHz~5.78GHz)

EFT bursts EN 61000-4-4:2012 Level 3 2KV/100KHz

Surge susceptibility EN 61000-4-5:2014 Level 4 2KV/Line-Line

Surge susceptibility EN 61000-4-5:2014 Level 4 4KV/Line-Earth

Conducted susceptibility EN 61000-4-6:2014 Level 3 10V

Magnetic field immunity EN 61000-4-8:2010 Level 4 30A/m

Voltage dip, interruption EN 61000-4-11:2004 100% dip 1 periods 30% dip 25 periods 100% interruptions 250 periods

#### Note:

A component power supply will be installed into final equipment. Since EMC performance will be affected by the complete installation, the final equipment manufacturers must re-qualify EMC Directive on the complete installation again.

For guidance on how to perform these EMC tests, please refer to TDF (Technical Documentation File)..

The product under declaration is just a unit without medical function. Complete MDR should only be verified when it is used together with particular medical device(s).

This Declaration is effective from serial number TC1xxxxxx

Person responsible for marking this declaration :

MEAN WELL Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/ Director, Group R&D :

(Name / Position)

(Signature)

Alex Tsai/Director, Product Strategy Center :

(Name / Position)

(Signature)

Taiwan

(Place)

May. 26th, 2021

(Date)