

DE 2-041736

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Medical Type Switching Power Supply		
Name and address of the applicant	MEAN WELL Enterprises Co., Ltd. No. 28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891 Taiwan		
Name and address of the manufacturer	MEAN WELL Enterprises Co., Ltd. No. 28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891 Taiwan		
Name and address of the factory	See additional page(s)		
Ratings and principal characteristics	AC 100-240V; 50/60Hz; 2.0-1.1A (for RPSy-160z); 1.7-0.9A (for RPDy-160z); 1.8-0.9A (for RPTy-160z and RPTy-160a-C); Class I; for output ratings refer to test report page 9 to 11		
Trademark (if any)	Trademark of MEAN WELL		
Customer's Testing Facility (CTF) Stage used	CTF Stage 1		
Model / Type Ref.	RPxy-160z; RPTy-160a-C (x = S, D or T for numbers of output; y = G or blank to denote different outputs; z = -5 , -12 , -15 , -24 , -48 , A, B, C or D to denote the differences in the output rating; a = A , B, C or D to denote the differences in the output rating		
Additional information (if necessary may also be reported on page 2)	Models differ in model name, output rating, main PCB layout, secondary circuit and layout; the risk management and usability requirements of the standard were not addressed; for details see test report		
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1+A2 for national differences see test report		
As shown in the Test Report Ref. No. which forms part of this Certificate	CN24VK33 001		

This CB Test Certificate is issued by the National Certification Body



Disclaimer: This is an electronically released document. The authenticity of this certificate can be verified on the IECEE Website "http://certificates.iecee.org"



DE 2-041736



10/061a CB 06/20v3 rk

			Page 2 01	2
1.	MEAN WELL Enterprise No. 28, Wuquan 3rd H Wugu District, New Taipei City 248 Taiwan	Rd.,		
2.	Suzhou MEAN WELL Teo No.269, Changping Ro Xiangcheng District, Jiangsu P.R. China	oad, Huangdai Town,		
3.			a, 2ND PHASE, PEENYA, 560058	
4.	MEAN WELL (GUANGZHOU CO., LTD. No. 11, Jingu South Road, Hu 510890 Guangzhou, P	uadu District,		
5.	Yongden Technology (345 MacArthur Highwa Guiguinto, Bulacan (Philippines	ay, Tabang,		
Additional information (if necessary) Report Ref. No. : CN24VK33 001				
			<u>a</u> 11-	
Date:	2024-04-24	Signature:	DiplIng. (FH) A. Klinke	er